




Idaho State Police Forensic Services

CERTIFICATE OF APPROVAL

The Idaho State Police Forensic Services (ISPFS) hereby certifies and approves Alcohol Simulator Solution Lot Number 10802 (a product manufactured by RepCo Marketing Inc.) to be used to conduct performance verification checks within the State of Idaho in accordance with the analytical methods, policies and/or procedures promulgated by the Department governing breath alcohol examinations. ISPFS also approves of the manufacturer of this solution (RepCo Marketing Inc.) to provide Alcohol Simulator Solution Lot Number 10802 in the State of Idaho. This lot has a target value of 0.080 with a range of 0.072 to 0.088 grams of ethyl alcohol/210 liters of vapor.


The expiration date for this lot number is August 9th, 2012 at 11:59 PM.

8-31-10
Date


Breath Alcohol Discipline Leader

STATE OF IDAHO)
) ss.
County of Kootenai)

On this 31st day of August, in the year 2010, before me, JoAnn Hutchison, a notary public, personally appeared Jeremy Johnston, known to me to be the person whose name is subscribed to the within instrument as a Forensic Scientist for the Idaho State Police Forensic Services, and acknowledge to me that he executed the same as such Scientist.

 Notary Public
My Commission Expires: 12/15/2015

